

# Certificate



Certificate no.

CU 72082147 01

**License Holder:**  
Crospon Ltd.  
Galway Business Park  
DANGAN, GALWAY  
Ireland

**Manufacturing Plant:**  
Crospon Ltd.  
Galway Business Park  
DANGAN, GALWAY  
Ireland

**Test report no.:** USA-GW 30892621 001

**Client Reference:** Adrian McHugh

**Tested to:** UL 60601-1:2003 R4.06  
CAN/CSA-C22.2 No. 601.1-M90

**Certified Product:** Functional Luminal Imaging Probe

**License Fee - Units**

**Model Designation:** EndoFLIP

7

**Rated Voltage:** DC 12V  
AC 100-240V, 47-63Hz (power adapter)  
**Rated Current:** 1.1-0.45A (power adapter)  
**Rated Power:** 80VA  
**Protection Class:** I, Type BF  
**Output Ratings DC:** 12V/6.6A (power adapter)

7

Appendix: 1, 1-2

**Licensed Test mark:**

**Signature**

**Date of Issue**

(day/mo/yr)

15/12/2008



*Martin Glagla*

**Dipl.-Ing. M. Glagla**  
**QA Certification Officer**